

<b>FREE LIQUIDS</b> <b>EPA 9095 B</b> <span style="float: right;">Page 1 of 1</span>					
Facility Name: _____ VELAP ID _____					
Assessor Name: _____ Analyst Name: _____ Inspection Date _____					
Relevant Aspect of Standards	Method Reference	Y	N	N/A	Comments
<i>Records Examined:</i> SOP Number/ Revision/ Date _____ Analyst: _____					
Sample ID: _____ Date of Sample Preparation: _____ Date of Analysis: _____					
Are conical paint filters with mesh number 60 +/- 5 % (fine meshed size) used? (Available at local paint stores.)	4.1				
Is the funnel of a size that 1 in of filter mesh protrudes?	4.2				
Is 100 gram or mL representative sample used?	6.0				
Note: materials such as sorbent pads or pillows should be cut into small pieces (< 1 cm).	7.3				
Note: Brittle materials > 1 cm may be lightly crushed or cut.	7.3				
Was sample allowed to drain for 5 minutes?	7.5				
If any portion of the material collects in the graduated cylinder was the material deemed to contain free liquids?	7.6				
Notes/ Comments:					